

GOODS OUT
MAR 23.

Serial No. C 103049

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

ENVIROCOM
CRANTHAM
LOCATION NCS1 7U1

Machine Details

Cust ref	PLATFORM
Make	A190E
Model	ERIC SCALE
Ser No	113650322
Capacity	1200 / kg
e/d	1 / 1 / kg
Working cap	1200 kg

Test Weights

Set No.	Cert No.	Date Cal
G SET	M1020E	09/01/23

Cert Interval 12 MONTHS.
Next Cert MARCH 2024

Tolerance +/-	kg
0- 500	1
500- 2000	2
2000 +	3

Comments
 ...MAKE UP LOAD USED...
 ...589 kg / 1060 kg...
 1 2
 3 4

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
100	100	0	100	0
200	200	0	200	0
539	539	0	539	0
739	737	-2	737	0
1060	1060	0	1060	0
1260	1257	-3	1257	0

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
1260	1257	1256	1257

Repeatability

Load	Ind	Ind	Ind

Eccentric Load

Reading	Dev	Reading	Dev
1 200.0	0	2 200	0
3 201	+1	4 200	0
5		6	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature

Name Shane Turner

Date 23/03/23