

STEP 2 - APR 23.

Serial No. C 103219

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

ENVIRONKOM LTD
SPIRE GATE LEVEL
GRANTHAM
NG31 7UH
LOCATION

Machine Details

Cust ref STEP 2
Make METTLER
Model IND560
Ser No 3073912
Capacity 21000/kg
e/d 10.5kg
Working cap

Test Weights

Set No.	Cert No.	Date Cal
9	09862	09/01/23

Cert Interval ANNUALLY
Next Cert APR 2024

Tolerance +/-	kg
0-500	0.5
500-1500	1
1500+	1.5

Comments
MAKE UP WEIGHT USED
692.5 kg

2	3
1	4

*

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
60	60	0	60	0
100	100	0	100	0
140	140	0	140	0
200	200	0	200	0
692.5	692.5	0	692.5	0
892.5	892.5	0	892.5	0

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability 692.5

Load	Ind	Ind	Ind
692.5	692.5	692.5	✓

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 200

Reading	Dev	Reading	Dev
1 200	0	2 200	0
3 200	0	4 200	0
5		6	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature Name J PEAKE Date 05/04/2023